



Product Code

- UTCM-0451 Vicat Apparatus
- UTCM-0452E Vicat Mould EN Base ID 80 mm,
Top ID 70 mm, Height 40 mm
- UTCM-0450A Vicat Mould ASTM Base ID 70 mm,
Top ID 60 mm, Height 40 mm
- UTCM-0453E Initial Vicat Needle EN 1.13 mm dia.
- UTCM-0453A Vicat Needle ASTM 1 mm dia.
- UTCM-0454E Final Vicat Needle EN 1.13 mm dia.
- UTCM-0455 Consistency Plunger 10 mm dia.
- UTCM-0458E Transfer Dish for UTCM-0451
- UTGT-1305 Glass Thermometer Max. 110°C
- UTCM-0457 Additiniol Weight, 700 gr, EN 480-2



Standards

EN 196-3, 480-2; ASTM C187, C191; AASHTO T129, T131


The Vicat test equipment is used for determining the setting time and consistency of cement by Vicat Method.

Vicat Apparatus consist of metal frame with a scale, Ø10mm consistency plunger and supporting glass plate.

The needle/s, mould, thermometer and transfer dish (for EN) should be ordered seperately.




Dimensions	150x190x318 mm
Weight (approx.)	3 kg




UTCM-0453A



UTCM-0455



UTCM-0453E



UTCM-0454E